

Instructions For Use

003

Accessory for urologic procedures

TF-15.02-003



1 Contents

2	Manufacturer information	3
3	Explanation of Symbols	3
4	General description.....	5
4.1	General warnings and cautions.....	5
4.2	Variants covered	5
5	Intended use.....	6
5.1	Intended purpose.....	6
5.2	Intended conditions of use	6
5.3	Intended use environment	6
5.4	Intended part of the body or type of tissue applied to or interacted with	7
5.5	Intended user profile	7
5.6	Intended medical indication	7
5.7	Intended patient population.....	7
6	Principles of operation and compatibility	7
7	Identification of parts	7
8	Operating Instructions	9
8.1	Priming	9
8.2	Spike.....	9
8.3	Roller Clamps	9
8.4	Luer locks	9
8.5	Flexible drip chamber.....	10
8.6	Check valve.....	10
8.7	Four-way tap / Stopcock	10
8.8	Caps.....	10
9	Cleaning / Disinfection / Disposal	10
9.1	Sterilization instructions	10

2 Manufacturer information

Manufacturer name: Lettix b.v.

Telephone: +31 (0)55 522 46 65

Email: info@lettix.nl

Address: Prinsenweide 66, 7317 BC Apeldoorn, The Netherlands

Website: www.lettixbv.com



Any serious incident that has occurred in relation to this device should be reported to the manufacturer and the competent authority of the Member State in which the user and/or patient is established.

3 Explanation of Symbols

	Indicates a medical device that has been sterilized using ethylene oxide
	Indicates a single sterile barrier system
	Indicates a single sterile barrier system with protective packaging outside
	Indicates that the fluid path is non-pyrogenic, ISO 7000-2723
	Indicated product is not in a sterile state
	Indicates that a medical device that should not be used if the package has been damaged or opened and that the user should consult the instructions for use for additional information
	Indicates a medical device that is intended for one single use only. Do not re-use
	Indicates a medical device that is not to be re-sterilized, this device is for Single Use Only.
	Indicates the manufacturer's catalogue number so that the medical device can be identified
	Indicates the manufacturer's batch code so that the batch or lot can be identified

	Indicates the item is a medical device
	Indicates a carrier that contains unique device identifier information
	Indicates the medical device manufacturer
	Indicates the date when the medical device was manufactured
	Indicates the entity distributing the medical device into the locale
	Indicates the date after which the medical device is not to be used
	Product contains or presence of DEHP
	CE mark including the number of the notified body
	Consult instructions for use
	Indicates a medical device that needs to be protected from moisture
	Keep away from sunlight, ISO 7000-0624

	Indicates the temperature limits to which the medical device can be safely exposed. Temperature limit, ISO 7000-0632
	Indicates the range of humidity to which the medical device can be safely exposed. Humidity limitation, ISO 7000-2620

4 General description

“Accessory for urologic procedures” are non-active non-implantable devices channeling liquids into the body. This medical device has a risk classification **Is** and is a CE-marked device.

Connecting tubing sets for urology are Well Established Technologies without a specific medical benefit.

Performance of the device is defined by its ability to serve as a fluid conduit. **Safety** of the device relates to the ability to maintain a closed system (no leak or ingress).

There is no direct **clinical benefit** for Connecting tubing sets for urology. However, bladder infusion sets (including the tubing) are a prerequisite for bladder infusion therapy. Therefore, bladder infusion sets have an indirect clinical benefit by making bladder infusion therapy possible. Bladder infusion sets allow the patient to benefit from the bladder infusion therapy benefits.

The following **technical** claims have been stated:

- Sterile products are sold in packaging that maintains a Sterility Assurance Level (SAL) of 10^{-6}
- Product meets applicable parts of the ISO 8536 standard
- Products are manufactured using medical-grade materials
- Products do not contain natural rubber latex

4.1 General warnings and cautions

Guideline on phthalates: As phthalates being marked as substances of concern, it is advised to limit exposure of products marked containing phthalates, to patient groups considered particularly vulnerable or exposed to high levels of phthalates such as, and including; paediatric population, peripubertal males, pregnant women, breast-feeding women.

In the event of the packaging being: 1) damaged; 2) unintentionally opened before use; and 3) if the packaging is exposed to environmental conditions outside of those specified, the device is not to be used and disposed of.

Re-use of single use devices can lead to the risk of cross contamination between patients.

Re-sterilization of single use devices can lead to a damaged devices.

4.2 Variants covered

This instruction for use is applicable to the following product variants:

Variant number	Variant name	Sterile / non-sterile	Containing phthalates (DEHP)
9037182	T.U.R. systeem, (275cm), S	Sterile	Yes
9045152	Blaasspoelsysteem, S	Sterile	Yes
9119152	Cystoscopiesysteem, S	Sterile	Yes
9292142	Cystoscopiesysteem 011, S	Sterile	Yes
9325152	Cystoscopiesysteem 012, S	Sterile	No
9371152	T.U.R. systeem, S	Sterile	Yes
9372152	Spoelsysteem urologie, S	Sterile	No
9388152	Cystoscopiesysteem, S	Sterile	No
9389132	TUR-systeem, S	Sterile	Yes
9404152	Cystoscopiesysteem, S	Sterile	No
9524142	Cystoscopiesysteem, S	Sterile	No
9571192	Urologisch systeem voor ps, S	Sterile	No
9604152	Cystoscopiesysteem enkel, S	Sterile	No
9614142	Cystoscopieset LL 250cm, S	Sterile	No
9615842	Cystometrieset LL 150cm, S	Sterile	No
9616142	Spoelset, S	Sterile	No
9617162	Terugslagklep met spike, S	Sterile	No
9623142	Cystoscopiesysteem dr. kamer, S	Sterile	No
9657152	Cystoscopie systeem enkel 160cm, S	Sterile	No
9693152	Cystoscopie systeem enkel, S	Sterile	No
9701142	TUR systeem enkel, 225cm, S	Sterile	No
9747152	Urologisch spoelsysteem, S	Sterile	Yes
9748132	T.U.R. Systeem, S	Sterile	Yes
9781142	Cystoscopiesysteem L=260cm, S	Sterile	Yes
9782152	Cystoscopiesysteem L=160cm, S	Sterile	No
9792362	T.U.R. systeem dubbel, S	Sterile	No
9799152	Cystoscopie systeem enkel, S	Sterile	No
9804152	Cystoscopie systeem enkel, S	Sterile	No

5 Intended use

5.1 Intended purpose

- Accessory for urologic procedures.
- May not be connected to an active device of class IIa and higher.
- Not intended to be used with bodily fluids.

5.2 Intended conditions of use

Sterile, non-measuring.

5.3 Intended use environment

Hospital use.

5.4 Intended part of the body or type of tissue applied to or interacted with

Non-invasive use.

5.5 Intended user profile

The product is intended to be used by trained healthcare practitioners, such as doctors or nurses. This means the user has a high level of education and professional competence. No disabilities are expected.

5.6 Intended medical indication

Patients who need a urologic procedure.

5.7 Intended patient population

It is advised to limit exposure of products marked containing phthalates to patient groups considered particularly vulnerable or exposed to high levels of phthalates such as, and including: paediatric population, peripubertal males, pregnant women, breast-feeding women.


6 Principles of operation and compatibility

This is an accessory used during a urologic procedure with a non-active medical device. Liquids are transported via tubing by gravity. Adaptors are used to connect the tubing to a urologic tool like a Cystoscope (female luer lock) or Catheter (catheter connection or silicon tube connection) and on the other side of the tubing to a liquid bag (spike) or pump (female luer lock). Transfer can be controlled by a clamp or tap. Y-connector allows the user to connect multiple irrigation liquid bags to the system, to ensure quick replacement of bags.

7 Identification of parts

Note: depending on the configuration of the accessory, different parts below are included. Not all accessories contain all parts identified below.

	Spike
	Clamp: roller
	Clamp: multistep

	Clamp: open/close
	Connector: luer locks
	Connector: rotating
	Connector: catheter
	Connector: Y
	Connector: Y + septum
	Flexible drip chamber
	Check valve
	Funnel

	Two way tap/stopcock
	Caps/protector

8 Operating Instructions

Note: depending on the configuration of the accessory, different parts below are included. Not all accessories contain all parts.

8.1 Priming

Follow hygiene policy, safety practices and identification policy of the hospital. Do not allow the free end of the giving set to touch any non-sterile surface, whilst you are running fluids through the line.

8.2 Spike

- Use protective gloves to prevent irritation from the leakage onto the hand of the operator.
- Take the IV bag into non-dominant hand
- Remove the stopper from the spiking port of the IV bag.
- Remove the cap from the spike, do not touch the sterile sections or allow them to come into contact with any non-sterile areas.
- Penetrate the spike into the spiking port while twisting back and forwards the spike.
- Make sure the spike is inserted as far as it will go, up to the raised ridge.

8.3 Roller Clamps

The primary function of a roller clamp is to control the flow rate of the IV fluid. The roller has two directions:

- Fully rolled up, flow rate is at its maximum.
- Roller halfway, the flow rate increased when the rolled up, and when it is rolled down, the flow rate goes down.
- Fully rolled down, the IV tubing is compressed and the flow stops.

8.4 Luer locks

The primary function of the luer locks is to make a leak-free connecting between tubing systems and medical- and Laboratory devices or other tubing systems. Luer locks has male- and female-taper fittings. Carefully connect the luer locks by pushing the female luer lock straight into the male luer lock using a clockwise, twisting motion. Tug gently to

ensure the connection is secure. Visually inspect for damage before use. Do not over tighten- may cause cracking of the luer and make removal difficult.

8.5 Flexible drip chamber

During priming the system, filling the drip chamber prevents air from entering the tubing, the drip chamber allows gas to rise out from a fluid so that the air is not passed downstream.

Fill the drip chamber one-third to one-half full by gently squeezing the chamber.

The drip chamber also helps estimate the rate at which fluid is administered.

8.6 Check valve

Prevents backflow into infusion line. This valve is not operatable.

8.7 Four-way tap / Stopcock

The valve is used to control the flow of the liquid and stops the flow when closed fully.

A 4-way stopcock allows for 360 degrees of rotation and has the states (shown below) for each of the four available positions. A 3-way stopcock has only three positions and has the first three states shown below.

- In the first state, liquid flows between points A and B.
- In the second, it flows between points A and C.
- In the third, it flows between points B and C.
- In the fourth state (4-way only), it flows between all three points.

8.8 Caps

Caps protect Luer locks connectors and spikes for contamination. Remove caps only according hospital procedures.

9 Cleaning / Disinfection / Disposal

Always dispose of blood contaminated products in a manner consistent with established biohazard procedures.

Dispose according to applicable legislation and/or infection prevention standards for healthcare waste, to minimize potential risk of contamination, infection or pollution.

9.1 Sterilization instructions

Not applicable. Products are sold sterile already and are for single use only.